



# Test Monitoring Center

<https://www.astmtmc.org>

A Program of ASTM International

## ASTM Reference Testing Semi-Annual Report ASTM D6557 – BRT

October 2021

# BRT Test Activity\*

Test Status	Validity Code	Number of Tests
Acceptable Calibration Test	AC	186
Failed Calibration Test	OC	6
Operationally Invalid, by Lab	LC	2
Operationally Invalid Shakedown Run, by Lab	LN	2
<b>Total</b>		<b>196</b>

- 5 labs reported data

\*April 1, 2021 – September 30, 2021

**Test Monitoring Center**  
<https://www.astmtncc.org>



# BRT Failed Tests

Failed Parameter	Number of Tests
Severe Average Gray Value	0
Mild Average Gray Value	6
<b>Total</b>	<b>6</b>

**Test Monitoring Center**

<https://www.astmtncc.org>



A Program of ASTM International

# BRT Failed Tests by Lab

Failed Parameter	LTMS Lab					#
	A	B	D	G	L	
Severe Average Gray Value	0	0	0	0	0	0
Mild Average Gray Value	2	1	2	1	0	6
Total	2	1	2	1	0	6

**Test Monitoring Center**

<https://www.astmtnmc.org>



A Program of ASTM International

# BRT Lost Tests\*

Status	Cause	#
Invalid	Heater malfunction	1
Invalid	Acid pump or syringe malfunction	1
<b>Total</b>		<b>2</b>

\*Invalid and aborted calibration tests

**Test Monitoring Center**  
<https://www.astmtncc.org>



A Program of ASTM International

# BRT Lost Tests by Lab

Cause	LTMS Lab					#
	A	B	D	G	L	
Power Outage	0	0	0	1	0	1
Acid pump or syringe malfunction	1	0	0	0	0	1
Total	1	0	0	1	0	2

# BRT Test Severity

- ▶ Average gray value is on target

**Test Monitoring Center**  
<https://www.astmtncc.org>




A Program of ASTM International

## REFERENCE AVERAGE GRAY VALUE

CUSUM Severity Analysis


Historical Chart




## REFERENCE AVERAGE GRAY VALUE

CUSUM Severity Analysis

Chart of recent results



# BRT Precision Estimates



# Information Letters\*

Test	Date	IL	Topic
			No new information letters this period.

\*Available from TMC Website

**Test Monitoring Center**  
<https://www.astmtmc.org>



A Program of ASTM International

# Reference Oil Inventory Estimated Life

Oil	TMC Inventory (gallons)	Quantity Shipped in last 6 months (gallons)	Lab Inventory (samples)	Estimated Life
1006	31.2	0.4	46	5+ years
82-1	3.5	0.4	46	4 years
86	50.8	0.4	40	5+ years
87	94.7	0.4	45	5+ years

# Miscellaneous Information

- ▶ Available on TMC Website:
  - Oil Assignment Request and Test File Upload
  - Live Reference Test Data Bases
  - Surveillance Panel Meeting Minutes
- ▶ <https://www.astmtmc.org/>



**A Program of ASTM International**