



## Test Monitoring Center

@ Carnegie Mellon University  
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>  
412-365-1000

Diesel Fuel Lubricity  
Reference Targets ASTM D6079  
High Frequency Reciprocating Rig (HFRR) 60°C

<b>Fluid Name</b>	<b>Reference Period</b>	<b>Target</b>	<b>Lower Acceptance</b>	<b>Higher Acceptance</b>
Low (Fluid B Lot 19)	02/01/2019 - 01/31/2020	685	632	738
High (Fluid A Lot 19)		361	284	438