



Test Monitoring Center

@ Carnegie Mellon University
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>
412-365-1000

Diesel Fuel Lubricity
Reference Targets ASTM D6079
High Frequency Reciprocating Rig (HFRR) 60°C

Fluid Name	Reference Period	Target	Lower Acceptance	Higher Acceptance
Low (Fluid B Lot 20)	01/01/2020 - 03/31/2021	674	604	744
High (Fluid A Lot 20)		423	405	442