



Test Monitoring Center

@ Carnegie Mellon University
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>
412-365-1000

Diesel Fuel Lubricity
Reference Targets ASTM D6079
High Frequency Reciprocating Rig (HFRR) 60°C

Fluid Name	Reference Period	Target*	Lower Acceptance*	Higher Acceptance*
Low (Fluid B Lot 13)	09/01/13 - 08/31/14	580	530	630
High (Fluid A Lot 13)	09/01/13 - 08/31/14	420	380	450

*Corrected of 20140702