



Test Monitoring Center

@ Carnegie Mellon University
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>
412-365-1000

Diesel Fuel Lubricity
Reference Targets ASTM D6079
High Frequency Reciprocating Rig (HFRR) 60°C

Fluid Name	Reference Period	Target	Lower Acceptance	Higher Acceptance
Low (Fluid B Lot 18-1)	03/07/2018 - 03/06/2019	578	541	615
High (Fluid A Lot 18-1)		369	262	476