



Test Monitoring Center

@ Carnegie Mellon University
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>
412-365-1000

CERTIFICATE OF ANALYSIS

TMC Product ID: 58
Product Purpose: Reference Calibration Check Oil for ASTM D6417
Product Description: Fully Formulated Engine Oil
TMC Batch Number: 19981001
Certificate Issue Date: 2020-07-02

Performance Targets in D6417*

Oil Code	Parameter	n	Mean	sR	95% Acceptance Limits**	
					Lower	Upper
58	Area % Evaporation Loss	18	5.61	0.30	5.0	6.2

*As established by the ASTM D02.B0.07 Volatility Surveillance Panel; see Lubricant Test Monitoring System (LTMS) D6417 Section for more details.

**95% Acceptance Limits = Mean \pm (1.96 sR)

This product has no established expiration date, empirical or otherwise, for an unopened container.

Store reference oils under cover in locations where the ambient temperature is between -10 °C and +50 °C.

Brittany J. Pfleegor
Supervisor, Analytical Testing
ASTM Test Monitoring Center
Ph: 412-365-1013
bjp@astmtmc.cmu.edu